



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/759,071	01/20/2004	Peter Awakowicz	53055US	7176

23911 7590 02/07/2006

CROWELL & MORING LLP
INTELLECTUAL PROPERTY GROUP
P.O. BOX 14300
WASHINGTON, DC 20044-4300

EXAMINER

JASTRZAB, KRISANNE MARIE

ART UNIT	PAPER NUMBER
----------	--------------

1744

DATE MAILED: 02/07/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/759,071

Applicant(s)

AWAKOWICZ ET AL.

Examiner

Krisanne Jastrzab

Art Unit

1744

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 18 November 2005.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-18 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-18 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Rejections - 35 USC § 112

The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

Claims 1-18 are rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter which was not described in the specification in such a way as to reasonably convey to one skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the claimed invention. The newly added claim limitation of expanding and condensing the vapor within "several tenths of a second up to less than 10 seconds" is not properly supported in the original disclosure. There is no recitation supporting the top end of that range at "less than 10 seconds".

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

Art Unit: 1744

1. Determining the scope and contents of the prior art.
2. Ascertaining the differences between the prior art and the claims at issue.
3. Resolving the level of ordinary skill in the pertinent art.
4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

Claims 1-18 are rejected under 35 U.S.C. 103(a) as being unpatentable over Cummings et al., U.S. patent No. 4,952,370 in view of Wu et al., U.S. patent No., 6,572,819 B1.

Cummings et al., teach sterilization of the surfaces of a chamber wherein a combination of steam and hydrogen peroxide is created in a vaporizer, the combination is sent to the chamber to be sterilized and then condensed on the surfaces being treated. A vacuum is drawn to remove the condensate by evacuation. The vacuum is set such that the water vapor will be removed first to enhance contact of the hydrogen peroxide. The steps of the process are repeated with the introduction of the hydrogen peroxide/steam combination occurring in a plurality of injections. Cummings et al., further teach that the expansion and condensation of the vapor can occur within 60

Art Unit: 1744

seconds. See column 2, lines 40-53, column 3, lines 40-68, column 4, lines 45-62, column 5, lines 20-30, column 6, lines 1-5, 12-16, 20-25, 33-50 and 65-68, and column 7, lines 1-5.

Wu et al., teach the known and expected use of non-conductive, non-reactive materials which can withstand exposure to sterilants such as steam and hydrogen peroxide for the construction of sterilization components. See column 2, lines 20-25, column 4, lines 5-11, and column 5, lines 30-57.

It would have been well within the purview of one of ordinary skill in the art to construct the sterilizer components of Cummings et al., of the materials taught in Wu et al., because such materials clearly withstand all parameters of the sterilization process to promote efficient sterilization of the articles to be treated.

Double Patenting

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. See *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and, *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent is shown to be commonly owned with this application. See 37 CFR 1.130(b).

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Claims 1-18 are provisionally rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 1-16 of copending Application No. 10/363,546. Although the conflicting claims are not identical, they are not patentably distinct from each other because they are of the same inventive concept, namely, sterilization by the application of steam and hydrogen peroxide such that they condense onto the surface being sterilized, with removal thereof by drawing a vacuum, differentiated only by '546 claiming placement of the object in a known and well recognized sterilization wrap or bag.

This is a provisional obviousness-type double patenting rejection because the conflicting claims have not in fact been patented.

Claims 1-18 are provisionally rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 1-20 of copending Application No. 09941,925. Although the conflicting claims are not identical, they are not patentably distinct from each other because they are of the same inventive concept, namely, sterilization by the application of steam and hydrogen peroxide such that they condense onto the surface being sterilized, with removal thereof by drawing a vacuum, differentiated only in that '071 specifies that the sterilization chamber be constructed of non-heat conducting materials which is intrinsic to the process requiring condensation of the sterilant onto those surfaces.

This is a provisional obviousness-type double patenting rejection because the conflicting claims have not in fact been patented.

Claims 1-18 are provisionally rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 1-8 of copending Application No. 10/806,292. Although the conflicting claims are not identical, they are not patentably distinct from each other because they are of the same inventive concept, namely, sterilization by the application of steam and hydrogen peroxide such that they condense onto the surface being sterilized, with removal thereof by drawing a vacuum, differentiated only by '292 claiming pre-heating of surfaces in the treatment area, a well recognized step in processes employing the injection of a previously vaporized sterilant in order to ensure that the sterilant reaches the entire area before condensing. This is a provisional obviousness-type double patenting rejection because the conflicting claims have not in fact been patented.

Response to Arguments

Applicant's arguments filed 11/18/2005 have been fully considered but they are not persuasive. Applicant has argued that Cummings fails to reach the limitation of condensation within less than 10 seconds, however, as Applicant's original disclosure does not properly support this limitation, it has not been considered and carries no weight, thus the rejection over Cummings properly stands maintained.

Finally, Applicant argues that the obviousness-type double patenting rejections are not proper because none of the claims would literally infringe the others and the invention are patentably distinct, however, the Examiner would disagree for the reasons cited in the rejection and Applicant has not clearly provided any basis for a non-obvious patentable distinction between the claimed inventions.

Conclusion

This is a RCE of applicant's earlier Application No. 10/759,071. All claims are drawn to the same invention claimed in the earlier application and could have been finally rejected on the grounds and art of record in the next Office action if they had been entered in the earlier application. Accordingly, **THIS ACTION IS MADE FINAL** even though it is a first action in this case. See MPEP § 706.07(b). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).


A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no, however, event will the statutory period for reply expire later than **SIX MONTHS** from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Krisanne Jastrzab whose telephone number is 571-272-1279. The examiner can normally be reached on Mon.-Wed. 6:30am-4:00pm and alternate Fridays.

Art Unit: 1744

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Rick Crispino can be reached on 571-272-1226. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



Krisanne Jastrzab
Primary Examiner
Art Unit 1744

February 3, 2006